



Canada's particle accelerator centre

How nuclear physics can treat cancer

Radiotherapy at TRIUMF

Cornelia Hoehr

Scientist


Deputy Associate Laboratory Director | Life Sciences





Tourism Vancouver

- **TRIUMF** – Tri University Meson Facility,
- since 1968, now 20 member universities
- Canada's particle accelerator centre



Member Universities:

- University of Alberta
- University of British Columbia
- University of Calgary
- Carlton University
- University of Guelph
- University of Manitoba
- Université de Montréal
- Queen's University
- University of Regina
- Simon Fraser University
- University of Toronto
- University of Victoria
- York University



Associate Members:

- McGill University
- McMaster University
- University of Northern BC
- Saint Mary's University
- Université de Sherbrooke
- Western University
- University of Winnipeg


TRIUMF - nuclear physics lab.

Expertise in:


- Accelerator technology
- Accelerator operation
- Detectors
- Targets for isotope production
- Interaction of particles




Applicable to
medicine




Cancer occurrence




Cancer mortality





Cancer occurrence




Cancer mortality



- Surgery
- Chemotherapy
- Ionizing radiation








- DNA (Deoxyribonucleic acid): genetic instructions for development and functioning
- Cell needs information from DNA for survival

- Single helix break easy to repair
- Double helix break more difficult to repair
- Cell can not survive


- **Radiotherapy: as many double helix breaks in cancer cells as possible with as few double breaks as possible in healthy cells**









Linear Energy Transfer (LET):
Energy transferred (ionization,
secondary electrons) per unit
distance




RBE - Relative Biological Effectiveness








OER – Oxygen Enhancement Ratio

- High LET radiation – low OER
- OER for x rays = up to 4
- OER for ions = 1






- **Electrons** only used for surface tumours
- **X-rays** (6-18 MV) most commonly used radiotherapy, many techniques to spare healthy tissue (3D conformal beams, image guided delivery, real-time motion tracking etc.), compact and cost efficient
- **Protons** need 230 MeV accelerator for clinical use, large facility, expensive



Conventional dose rate ~ 0.03 Gy/s

- In 20 – 30 fractions to affect all cell cycle phases, and to reach the hypoxic centre of a tumour





Conventional dose rate ~ 0.03 Gy/s

- In 20 – 30 fractions to affect all cell cycle phases, and to reach the hypoxic centre of a tumour

FLASH dose rate < 40 Gy/s

- Lower toxicity in healthy tissue but same tumour control
- Effect only consistently observed *in-vivo*, not *in-vitro*
- Oxygen – depletion hypothesis, healthy tissue becomes basically hypoxic
- To reach high dose rates – remove target..... **electron beam**



Conventional dose rate ~ 0.03 Gy/s

- In 20 – 30 fractions to affect all cell cycle phases, and to reach the hypoxic centre of a tumour

FLASH dose rate < 40 Gy/s

- Lower toxicity in healthy tissue but same tumour control
- Effect only consistently observed *in-vivo*, not *in-vitro*
- Oxygen – depletion hypothesis, healthy tissue becomes basically hypoxic
- To reach high dose rates – remove target..... **electron beam**




1a : Day 0





1b : 3 weeks




1c : 5 months




- FLASH in Vancouver
- Around 260 Gy/s
- Data still being analyzed







- Electron gun 300 keV 10 mA (CW)
- Three accelerating superconductive cryomodules
- Irradiation stations:
 - Low energy (CTS – 300 keV)
 - High energy (EABD – up to 30 MeV)
 - Medium energy (EMBD - up to 10 MeV)



Magdalena Bazalova-Carter, Uvic
Alexander Gottberg, TRIUMF





- Electron gun 300 keV 10 mA (CW)
- Three accelerating superconductive cryomodules
- Two side irradiation stations:
 - Low energy (CTS – 300 keV)
 - High energy (EABD – up to 30 MeV)
 - Medium energy (EMBD - up to 10 MeV)




Magdalena Bazalova-Carter, Uvic
Alexander Gottberg, TRIUMF

FLASH – DoseEq @ 10 MeV

 Average dose rate up to ~ 300 Gy/s


What's next?

- Manufacture, installation and testing until summer 2020
- Experiments with biological samples fall 2020






Hans
Bethe

$$-\frac{dT}{dx} = \frac{4\pi e^4 z^2}{m v^2} Z \ln \frac{2 m v^2}{E},$$

Zur Theorie des Durchgangs schneller Korpuskularstrahlen durch Materie, Annalen der Physik. vol. 397, pp. 325-400, 1930




Robert
Wilson

Radiological Use of Fast Protons, Radiology vol. 47, pp. 487-91, 1946






Proton and heavy ion therapy centers



Currently 91 centers in operation, 46 under construction, 21 in planning

- Clinical operation since 1995
- Canada's only proton therapy facility
- Ocular melanomas
- Clinical treatments ended Feb 2019






Summary paper with 59 patients

(E. Tran et al., Int. J. Radiat. Oncol. Biol. Phys. 83 (2012) 1425)



- 20 patients T1, 28 patients T2, 11 patients T3
- Median tumor size: diameter 11.4 mm, 3.5 mm thick
- Median follow-up time 63 month
- 19 patients treated with 54 CGE and 40 patients treated with 60 CGE
- **5-year local control rate 91%**
(T1 100%, T2 93%, T3 59%) **and 97%** with 60 CGE, 83% 54 CGE
- **Metastasis-free survival rate 82%** (T1 94%, T2 84%, T3 47%)
- 5-year neovascular glaucoma 31% (T1-2 23%, T3 68%)
- Enucleation T1 0%, T2 14%, T3 72%





before PT




after PT





- Typical sensors for dosimetry larger than treatment volume
- Optical fibers can have excellent spatial resolution
- Dose and dose rate independent








$$\frac{dL}{dx} = \frac{S \frac{dE}{dx}}{1 + k_B \frac{dE}{dx} + C \left(\frac{dE}{dx} \right)^2}$$


- Empirical Craun-Birks equation
- Correction of quenching
- Not practical for Proton Therapy!































- Summer 2019: experiment with HPGe detector
- Data currently being analyzed by Eva and Christina






- PT facility (2C1) at TRIUMF limited to 6 nA, or ~ 0.2 Gy/s
- Main cyclotron able to extract 100 μ A into 2C4, or up to 3,333 Gy/s
- Will there be a FLASH effect with protons?
- How do you ensure range verification?

Biologically effective doses for photons, protons and carbon ions





Kratochwil *et al.*, *J. Nuc. Med.* 2016;57(12):1941–1944.



- 11 clinical trials (^{225}Ac and ^{213}Bi)
- > 640 patients (60-80% showed response)
- Want up to 50,000 patient doses a year (120 Ci)

Kratochwil *et al.*, *J. Nuc. Med.* 2016;57(12):1941–1944.

Primary ^{225}Ac sources:


- $^{229}\text{Th}/^{225}\text{Ac}$ generator ($t_{1/2} \sim 7880$ y) sourced via legacy stockpile, ORNL, ITU
- Alternatives sought
- ^{226}Ra irradiation
- Tri-Lab efforts $^{232}\text{Th}(p,x)$ spallation

Global production is ~ 10 Ci per year


- Promising early clinical trial results
- Supply vs demand is out of balance, but market needs to be nurtured, and supply needs to increase and be reliable

- ^{225}Ac production via Th spallation demonstrated at small scales:

LANL	50 - 200, 800 MeV
INR RAS	40 - 90 MeV




J.W. Weidner et al. Appl. Radiat. Isotop. 2012, 70, 2602



Isotope production using
TRIUMF's
500 MeV infrastructure



IPF (BL1A)
Intermediate activity (MBq),
spallation


- Routine, independent
production



500 MeV Isotope Production Facility



- Hundreds of co-produced isotopes including
- ^{225}Ra , ^{225}Ac , ^{224}Ra , ^{223}Ra , ^{213}Bi , ^{212}Pb , ^{212}Bi , $^{209/211}\text{At}$








concerns (from some) about ^{227}Ac content and impact on waste management — no consensus

directly-produced ^{225}Ac



concerns (from some) about ^{227}Ac content and impact on waste management — no consensus



directly-produced ^{225}Ac



^{225}Ac + ^{224}Ac + ^{226}Ac + ^{227}Ac
 $t_{1/2} = 9.9 \text{ d}$ $t_{1/2} = 3 \text{ h}$ $t_{1/2} = 29 \text{ h}$ $t_{1/2} = 22 \text{ y}$

concerns (from some) about ^{227}Ac content and impact on waste management

directly-produced ^{225}Ac



^{225}Ac + ^{224}Ac + ^{226}Ac + ^{227}Ac
 $t_{1/2} = 9.9 \text{ d}$ $t_{1/2} = 3 \text{ h}$ $t_{1/2} = 29 \text{ h}$ $t_{1/2} = 22 \text{ y}$

concerns (from some) about ^{227}Ac content and impact on waste management

directly-produced ^{225}Ac

^{225}Ac + ^{228}Ac
 $t_{1/2} = 9.9 \text{ d}$ $t_{1/2} = 6 \text{ h}$


generator-produced ^{225}Ac

Process steps:







- 1) Th Irradiation
- 2) Ra/Ac separation 1 week EOB – gives primary Ac fraction
- 3) Recovered Ra allowed to sit for 17.5 days
- 4) Ra/Ac separation – gives secondary Ac fraction

Primary Ac Fraction			
days from Ra/Ac isolation	0	1	5
Ac-225 [MBq]	42.2	39.4	29.9
Ac-228/Ac-225 [%]	0.039	0.003	0.000
Ac-227/Ac-225 [%]	0.185	0.198	0.261
Ac-226/Ac-225 [%]	16.020	9.740	1.330

Secondary Ac Fraction		
days from Ra/Ac isolation	0	2
Ac-225 [MBq]	2.2	1.9
Ac-228/Ac-225 [%]	0.882	0.003
Ac-227/Ac-225 [%]	9.951E-09	9.949E-09
Ac-226/Ac-225 [%]	0.000	0.000



- Up to 6 cassettes, 12 targets
- Targets immersed in circulating water bath
- Thermocouples monitor target, water temperatures
- Cassettes can be moved without beam on/off



Irradiation parameters

- integrated current 2640 $\mu\text{A}\cdot\text{h}$, over **36-40 h**
- 66-73 μA , 454 MeV


Results (n = 3)


- (521 ± 18) MBq ^{225}Ac and (95 ± 24) MBq ^{225}Ra at EOB
- saturation yields 72 MBq/ μA (^{225}Ac), 19 MBq/ μA (^{225}Ra)
- (319 ± 11) MBq ^{225}Ac and (69 ± 17) MBq ^{225}Ra at transfer to radiochemistry



What's next?

- **x12 increase** in yield by irradiating 12 targets
- **x10 increase** in yield by irradiating for full ^{225}Ra half life (15 days)
- Further increase from thicker target and higher current require re-evaluation of target and safety







Isotope production using
TRIUMF's
500 MeV infrastructure


ARIEL/ H^+

High activity (GBq), spallation

- Enable radiopharmaceutical development and clinical trials

- 400 mCi (15 GBq) ^{225}Ac per target (FLUKA)
- Irradiation schedule decoupled from science target
- ARIEL Proton Station commissioning scheduled for 2024/5 – 10 M CFI grant





TRIUMF Life Sciences Division

TRIUMF accelerator division,
operations and machine shops

Collaborators

- Cheryl Duzenli (BC Cancer)
- Francois Bénard (BC Cancer)
- Andrew Minchinton (BC Cancer)
- Andrew Jirasek (UBC)
- Boris Stoeber (UBC)
- Dennis Muecher (Guelph)
- Magdalena Bazalova-Carter (UVic)
- Sylvain Girard (St Etienne, France)
- Sinead O’Keeffe (Limerick, Ireland)



New Frontiers in Research Fund
Fonds Nouvelles frontières en recherche



Canada



CIHR IRSC
Canadian Institutes of Health Research
Instituts de recherche en santé du Canada





Canada's particle
accelerator centre


Centre canadien
d'accélération des
particules

TRIUMF: Alberta | British Columbia | Calgary |
Carleton | Guelph | McGill | Manitoba | McMaster |
Montréal | Northern British Columbia | Queen's |
Regina | Saint Mary's | Simon Fraser | Toronto |
Victoria | Western | Winnipeg | York



Thank you!
Merci!


Follow us at TRIUMFLab





- Pion - subatomic particle, meson
- In nuclei, glue to hold protons and neutrons
- Some are charged
- Have Bragg peak, little damage to surrounding tissue, high LET in Bragg peak
- Lots of damage at Bragg peak ('pion star')





- Study from 1980 – 1994 (over 300 patients), one of only three in the world
- Brain tumors (glioblastoma) and prostate cancer
- Result of study: no advantage over conventional photon therapy




Fig. 2. Overall survival for both treatment groups. Median survivals are: photons, 10 months; pions, 10 months. Log rank: $p = 0.22$.